


**ATTACHMENT E**

**LOW/HIGH TEMPERATURE THERMAL CIRCULAR CHARTS**

**This page intentionally left blank.**



WYLE LABS	J/NR	45827	
	CUSTOMER	NOTONIC	
	TYPE TEST	High Temp Sp.	
	DRY BULB	WET BULB	
	CHAMBER		
	START DATE	2-2-01	
TECHNICIAN	C. J. [unclear]		
CHECKED BY		DATE	2-2-01



**ATTACHMENT F**  
**VIBRATION TEST DATA**

This page intentionally left blank.

# VIBRATION TEST DATA SHEET

Customer ESS Spec. Mic. 570-810D Specimen DRE COILING MACHINE  
 Job No. 45827 Method S14.1 Part No. VIBRONIC 2000 Specimen Temp. AMB.  
 GSI Yes  No  Procedure --- S/N 000003 - VeriVibe Photo Yes  No   
2001B (-Comm) Pice


Test Title TRANSIT VIBRATION

Date	Time	Axis	Temp (F)	SINUSOIDAL			RANDOM			TOTAL Accel. (grms)	Test Time (min)	COMMENTS	NAME
				Freq. (cps)	Disp. (in)	Accel. (±g)	Freq. (cps)	PSD (g <sup>2</sup> /cps)	Slope (dB/oct)				
5/1/01	1157	TRANS VERT	AMB				10-40	.015			Run #1 2 specimens	April	
							500	.00015	1.04	30			
5/1/01	1231	AMB	AMB				10-40	.015			Run #2 2 specimens	April	
							500	.00015	1.04	30			
5/1/01	1335	VERT	AMB				10-40	.015			Run #3 SOFT CASE ONLY	April	
							500	.00015	1.03	30			
5/1/01	1412	VERT	AMB				10-40	.015			Run #4 HARD CASE ONLY		
							500	.00015	1.04	30			


Job No. 45827  
 Report No. 45827-01  
 Date 5/1/01  
 Page 1 of 1

WH-1028A  
 Signed [Signature] 5/1/01  
 Approved [Signature] 5/1/01


Run#1 CONTROL AVG. 2 SPECI. NS 5/1/01  
xdc:sa POST TEST ELAPSED TIME = 1802 SECS AT .00 DB  
RMS LEVEL = 1.041 G'S DELTA F = 2.000 DOF = 180 AUF = 8




RUN#2 CONTROL AVG. 2 SPECI \_NS 5/1/01  
POST TEST ELAPSED TIME - 1802 SECS AT .00 DB  
RMS LEVEL - 1.045 G'S DELTA F - 2.000 DOF - 180 ALF - 8



FLINCH CONTROL AUG. SOFT CAL ONLY 5/1/01  
POST TEST  
ELAPSED TIME = 1804 SECS AT .00 DB  
RMS LEVEL = 1.039 G'S DELTA F = 2.000 DOF = 180 AUF = 8



kdci:sa RUN#4 CONTROL AUG. HARD CAL ONLY 5/1/01  
POST TEST ELAPSED TIME = 1804 SECS AT .00 DB  
RMS LEVEL = 1.041 G'S DELTA F = 2.000 DOF = 180 AUF = 8



**This page intentionally left blank.**

**ATTACHMENT G**  
**BENCH HANDLING DATA SHEET**

This page intentionally left blank.

### DATA SHEET

Customer ES&S WYLE LABORATORIES  
Specimen DRE Voting Machine  
Part No. iVotronic 2000 Amb. Temp. ~74°F Job No. 45587  
Spec. FEC Standard Photo - Report No. 45587-01  
Para. 7.3.2.3 Test Med. - Start Date 5.7.01  
S/N 10065 Specimen Temp. ~74°F  
GSI No  
Test Title Bench Handling

Criteria: Using one edge (base of machine) as a pivot, the opposite edge shall be raised to a height of four inches above the surface and allowed to drop freely for six drops. Each of the remaining base edges shall be similarly dropped.

Edge 1: Drops 1 - 6

Edge 2: Drops 7 - 12

Edge 3: Drops 13 - 18

Edge 4: Drops 19 - 24

Post-Test Inspection: No degradation/damage observed. Unit remained fully functional.

Tested By *[Signature]* Date: 5/7/01  
Witness - Date: -  
Sheet No. 1 of 1  
Approved *[Signature]* S.L.V.

Notice of Anomaly: -

Wyle Form WH 614A. Rev. APR '84


**This page intentionally left blank.**


ATTACHMENT H  
24-HOUR HUMIDITY THERMAL CIRCULAR CHARTS


**This page intentionally left blank.**


14-HOORVIMUITY THE NAME JHT YTKIMBUA800K-42


























**This page intentionally left blank.**

ATTACHMENT I

DUST TEST DATA SHEET AND THERMAL CIRCULAR CHARTS

This page intentionally left blank.

DATE TESTED: 04/24/07 BY: JEFFREY WATSON AT: 12800

## DATA SHEET


Customer	<u>ES&amp;S</u>	<b>WYLE LABORATORIES</b>
Specimen	<u>DRE Voting Machine</u>	
Part No.	<u>iVotronic 2000</u>	Amb. Temp. <u>~74°F</u>
Spec.	<u>FEC Standard</u>	Photo <u>--</u>
Para.	<u>7.3.2.9</u>	Test Med. <u>--</u>
S/N	<u>00100063</u>	Specimen Temp. <u>~74°F</u>
GSI	<u>No</u>	
Test Title	<u>Sand &amp; Dust Exposure</u>	

<p>Sand &amp; Dust Exposure Testing shall be performed on the iVotronic while stowed within a black nylon canvas carrying case and also with the iVotronic stowed within an ES&amp;S provided voting booth. Additionally, the Communications Pack (S/N 200089) shall be tested while stowed in a hard-shell carrying case.</p>
<p>Each unit was subjected to the Sand &amp; Dust exposure in accordance with Mil-Std-810D, Section II-1.1.1.</p>
<p>iVotronic Post-test inspection (Canvas Carrying Case): Very minor traces of dust intrusion were observed within the iVotronic carrying case. Continued post-test functionality was verified. The iVotronic remained fully functional.</p>
<p>iVotronic Post-test inspection (Voting Booth): Very minor traces of dust intrusion were observed within the iVotronic voting booth. Continued post-test functionality was verified. The iVotronic remained fully functional.</p>
<p>Communications Pack (Hard-shell Case): Dust intrusion into the hard-shell case was observed. The dust was blown clear of the printer. The printer was powered and remained fully functional.</p>


Tested By	<u><i>[Signature]</i></u>	Date:	<u>10-3-01</u>
Witness	<u>-</u>	Date:	<u>-</u>
Sheet No.	<u>1</u>	of	<u>1</u>
Approved	<u><i>[Signature]</i></u>		<u>10.3.01</u>


Notice of Anomaly: -

WYLE LABS	JVNR	45827
CUSTOMER	255	
TYPE TEST	DUST	
DRY BULB	2	WET BULB 3
CHAMBER	29	
START DATE	8-9-01	
TECHNICIAN	[Signature]	
CHECKED BY:	[Signature]	
DATE	8-20-01	



NO	JANR	45827
Q	CUSTOMER	ESS
A	TYPE TEST	DST
L	DRY BULB	29
W	CHAMBER	3-10-01
W	START DATE	12-1-01
W	TECHNICIAN	
	CHECKED BY	
		DATE 12-20





**This page intentionally left blank.**

**ATTACHMENT J**  
**RAIN (DRIP) TEST DATA SHEET**

**This page intentionally left blank.**

**DATA SHEET**

Customer	<u>ES&amp;S</u>			WYLE LABORATORIES
Specimen	<u>DRE Voting Machine</u>			
Part No.	<u>iVotronic 2000</u>	Amb. Temp.	<u>~74°F</u>	Job No. <u>45827</u>
Spec.	<u>FEC Standard</u>	Photo	<u>--</u>	Report No. <u>45827-01</u>
Para.	<u>7.3.2.8</u>	Test Med.	<u>--</u>	Start Date <u>5.10.01</u>
S/N	<u>00100063</u>	Specimen Temp.	<u>~74°F</u>	
GSI	<u>No</u>			
Test Title	<u>Rain (Drip) Exposure</u>			

<p>Rain (Drip) Testing shall be performed on the iVotronic while stowed within a black nylon canvas carrying case and also with the iVotronic stowed within an ES&amp;S provided voting booth. Additionally, the Communications Pack (S/N 200089) shall be tested while stowed in a hard-shell carrying case.</p>
<p>Each unit was subjected to a drip rate of 7 gallons/sq.ft./hr for 15 minutes with a drip height of 3 feet. Upon completion, each unit shall be inspected for obvious signs of water intrusion.</p>
<p>iVotronic Post-test inspection (Canvas Carrying Case): Slight traces of water intrusion were observed within the iVotronic carrying case. However, the iVotronic remained fully functional. Continued post-test functionality was verified. There was no visible evidence to suggest that any water had penetrated the iVotronic.</p>
<p>iVotronic Post-test inspection (Voting Booth): Slight traces of water intrusion were observed within the iVotronic voting booth. However, the iVotronic remained fully functional. Continued post-test functionality was verified. There was no visible evidence to suggest that any water had penetrated the iVotronic.</p>
<p>Communications Pack (Hard-shell Case): Some water intrusion was present within the hard-shell case and on the printer. The printer was wiped and powered and remained fully functional.</p>

Tested By Jan Larson Date: 10-3-01  
 Witness - Date: -  
 Sheet No. 1 of 1  
 Approved Jan Larson

Notice of Anomaly: -

Wyle Form WH 814A. Rev. APR '84

This page intentionally left blank.

**ATTACHMENT K**  
**INSTRUMENTATION EQUIPMENT SHEETS**

**This page intentionally left blank.**



INSTRUMENTATION EQUIPMENT SHEET

DATE: 5/1/01      JOB NUMBER: 45827      TEST AREA: DYN LAB  
 TECHNICIAN: L.MATHIS      CUSTOMER: ESS      TYPE TEST: VIB

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	OSCOPE	TEKTRONIX	2213	N/A	101036	60MHZ	4%	4/10/01	10/5/01
2	HARD COPY UNIT	TEKTRONIX	4631	B187407	113330	MULTI	MFG	3/23/01	9/19/01
3	TERMINAL	TEKTRONIX	4612	N/A	100589	MULTI	MFG	1/16/01	1/16/02
4	VIB CONTROL	HP	5427A366	2120A00409	100291	MULTI	MFG	11/22/00	11/22/01
5	DMM	KEITHLEY	179A	196804	101203	MULTI	MFG	9/28/00	9/28/01
6	OSCOPE	TEKTRONIX	2213A	B013724	101481	60MHZ	3%	4/10/01	10/5/01
7	ACCEL	BRUEL & KJAER	4366	1104817	101774	2KGSV/SKGSK	5%	4/16/01	7/16/01
8	CHARGE AMP	ENDEVCO	2775A	ED75	112653	GAIN	1.5%	1/31/01	7/30/01
9	ACCEL	BRUEL & KJAER	4366	1104831	101814	2KGSV/SKGSK	5%	4/16/01	7/16/01
10	CHARGE AMP	ENDEVCO	2775A	EE30	112651	GAIN	1.5%	1/31/01	7/30/01

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION [Signature] 5/1/01      CHECKED & RECEIVED BY [Signature] 5-1-01

Q.A. Banda Marco 5-11-01

WH-1029A, REV. APR '99



INSTRUMENTATION EQUIPMENT SHEET

DATE: 5/1/01      JOB NUMBER: 45827      TEST AREA: ENV CH 50  
 TECHNICIAN: LIVEY      CUSTOMER: VOTRONIC      TYPE TEST: TEMPERATURE

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	TEMP RECORDER	HONEYWELL	DR4500	9634Y6268573	112982	-200-600°F	.4°F	3/7/01	6/5/01
2	TEMP CONTROLLE	THERMOTRON	6800	N/A	105286	-148 to 437°F	±.25%	12/4/00	6/1/01
3	TEMP ALARM	THERMOTRON	12005	263002	094751	-125-375°F	.25%	12/4/00	6/1/01

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION Sammy S-C-01 CHECKED & RECEIVED BY [Signature] 5.11.01  
 Q.A. [Signature] 5/1/01

WH-1029A, REV. APR '99



INSTRUMENTATION EQUIPMENT SHEET

DATE: 5/2/01      JOB NUMBER: 45827      TEST AREA: ENV CH 7  
 TECHNICIAN: LIVEY      CUSTOMER: VOTRONIC/ESS      TYPE TEST: HUMIDITY

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	TEMP ALARM	THERMOTRON	012005	N/A	092490	-125 to 375°F	.25%	3/16/01	9/12/01
2	MICROCOMPUTER	THERMOTRON	3038	7626-TP	114130	-125 - 375°F	.25%	3/16/01	9/12/01
3	TEMP RECORDER	HONEYWELL	DR4500	9628Y6262947	112722	-200-600°F	.4°F	3/16/01	6/14/01

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION

*Laurel 5-3-01*

CHECKED & RECEIVED BY

*[Signature]* S.3.0  
 Q.A. *[Signature]* 5-3-01

WH-1029A, REV. APR 79



INSTRUMENTATION EQUIPMENT SHEET

DATE: 5/14/01      JOB NUMBER: 45827      TEST AREA: ENV CH 51  
TECHNICIAN: J.LAXSON      CUSTOMER: ESS      TYPE TEST: TEMP CYCLING

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	DMM	FLUKE	87	60530991	112224	MULTI	MFG	4/ 3/01	4/ 3/02
2	ALARM LIMIT	NEWPORT	Q2001TC	N/A	105433	-184 -- -59°C	1.5%FS	12/ 7/00	6/ 5/01
3	TEMP RECORDER	HONEYWELL	DR450T	924488505000	109830	-200-600°F	4°F	3/30/01	6/28/01
4	TEMP CONTROLLE	RESEARCH	828-B11	10033	108416	-400-700°F	.1%FS	12/ 7/00	6/ 5/01

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION J. A. Larson 5-14-01 CHECKED & RECEIVED BY J. Larson S.H.  
Q.A. Bonda Mac 5.14.01

WH-1029A, REV. APR 99



INSTRUMENTATION EQUIPMENT SHEET

DATE: 5/18/01      JOB NUMBER: 45827.00      TEST AREA: PACK TECH  
TECHNICIAN: T.BATES      CUSTOMER: ESS      TYPE TEST: DROP \*

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	SCALE	OHAUS	CD11	00082316GA	116051	0-50LBS	+/-0.1 LBS	3/14/01	3/14/02
2	SCALE	HOWE RICHARDSON	5402	79-08130	101917	2KLB	.5LB	10/27/00	10/26/01

\* Fiberboard Pallet

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION

5-21-01

CHECKED & RECEIVED BY

5-21-01

P.A. Brenda Moore 5-21-01

WH-1029A, REV. APR '99



INSTRUMENTATION EQUIPMENT SHEET

DATE: 8/9/01  
 TECHNICIAN: L.IVEY

JOB NUMBER: 45827  
 CUSTOMER: BSS

TEST AREA: ENV CH 29  
 TYPE TEST: DUST

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	CONTR DUST	PHOTOMATION	DSMIPB	246438	101480	5G/CU FT 0-10	2%	5/1/01	5/1/02
2	HUM XMITTER	HICAL	CT-829AR	112055	112055	20-80%RH	MFG	7/17/01	1/11/02
3	TEMP ALARM	RESEARCH	61034	31524-01-49	094325	-175 - 375°F	±0.5%	4/2/01	9/28/01
4	TEMP CONTROLLE	RESEARCH	61011		000717	-175-325°F	±.5%	4/18/01	10/15/01
5	TEMP RECORDER	HONEYWELL	DR450T	920284912600	109510	-200-600°F	.4°F	8/9/01	11/7/01

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION

*Troy Long* 8-9-01

CHECKED & RECEIVED BY

*[Signature]* 8-9-01

Q.A.

*Burda* *Mason* 8-9-01



INSTRUMENTATION EQUIPMENT SHEET

DATE: 8/10/01  
TECHNICIAN: T.BATES

JOB NUMBER: 45827  
CUSTOMER: ESS

TEST AREA: PACK TECH  
TYPE TEST: DROP #

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	SCALE	OHAUS	CD11	00082316GA	116051	0-50LBS	±.01 LBS	3/14/01	3/14/02

\* Nylon carrying bag ;  
Using booth

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION

*Titus Bates 8/13/01*

CHECKED & RECEIVED BY

*[Signature] 8/13/01*

Q.A.

*[Signature] 8/13/2001*

WH-1029A, REV, APR 99



INSTRUMENTATION EQUIPMENT SHEET

DATE: 8/10/01      JOB NUMBER: 45872      TEST AREA: ENV. LAB  
TECHNICIAN: J.MCDERMOTT      CUSTOMER: EGS      TYPE TEST: RAIN - DRIP

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	STOP WATCH	CRONUS	603	NA	115808	10HR	0.5 SEC	6/29/01	9/27/01

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION Langley 8-10-01 CHECKED & RECEIVED BY [Signature]  
o.a. Brenda Morse 8.10.01

WH-1029A, REV. APR 99

**ATTACHMENT L**  
**IVOTRONIC PRODUCT SAFETY REPORT**

**This page intentionally left blank.**

WYLE LABORATORIES, INC.  
Page 1

Report No.: 45827-02  
E S & S


Issued: 9/11/2001

REPORT NO. 45827-02

INSPECTION, TEST AND EVALUATION  
OF THE  
IVOTRONIC VOTING MACHINE

SUBMITTED TO

ELECTION SYSTEMS AND SOFTWARE  
11208 JOHN GALT BOULEVARD  
OMAHA, NEBRASKA 68137 - 2364



**WYLE LABORATORIES, INC.**

WYLE is a Nationally Recognized Testing Laboratory (NRTL)

COPYRIGHT BY WYLE LABORATORIES. THE RIGHT TO REPRODUCE, COPY, EXHIBIT, OR OTHERWISE UTILIZE ANY OF THE MATERIAL CONTAINED HEREIN WITHOUT THE EXPRESS PRIOR PERMISSION OF WYLE LABORATORIES IS PROHIBITED. THE ACCEPTANCE OF A PURCHASE ORDER IN CONNECTION WITH THE MATERIAL CONTAINED HEREIN SHALL BE EQUIVALENT TO EXPRESS PRIOR PERMISSION.

WYLE LABORATORIES, INC.  
Page 2

Report No.: 45827-02  
ES & S

Issued: 9/11/2001

**LISTING REPORT**  
**WYLE LABORATORIES, INC.**

7800 Highway 20 West, PO Box 077777, Huntsville, Alabama 35807

Purchase Order No. 512395-00

REPORT NO. 45827-02  
INSPECTION, TEST AND EVALUATION  
OF THE

**IVOTRONIC VOTING MACHINE**

SUBMITTED TO  
ELECTION SYSTEMS AND SOFTWARE  
11208 JOHN GALT BOULEVARD  
OMAHA, NEBRASKA 68137-2364

**GENERAL:** This Report gives the results of the inspection, test and evaluation of the IVOTRONIC VOTING MACHINE for compliance with the applicable requirements of the "Standard for Safety of Information Technology Equipment," UL 60950, Third Edition. Mr. Tim Cordes authorized this investigation. Samples in good condition were provided by the client and tested at Wyle Labs' Huntsville facility.

Safety of Information Technology Equipment, UL 60950

**Applicant:** Election Systems and Software  
11208 John Galt Boulevard  
Omaha, Nebraska 68137-2364

**Contact:** Mr. Tim Cordes  
**Telephone:** (800) 247-8683  
**Fax:** (403) 593-8107

**Manufacturer:** Pivot International  
14125 West 95<sup>th</sup> Street  
Lenexa, Kansas 66215

**Contact:** Mr. Shawn Thompson  
**Telephone:** (913) 438-5210, ext. 1219  
**Fax:** (913) 438-5201



WYLE LABORATORIES, INC.

Page 4

Report No.: 45827-02  
E S & S

Issued: 9/11/2001

**PRODUCT DESCRIPTION**

**PRODUCTS COVERED:**

Ivotronic Voting Machine

**PRODUCT DESCRIPTION:**

The products covered by this report are an Ivotronic Voting Machine. The Ivotronic is a system used for collecting and tallying of votes. The total weight of the Ivotronic Voting Machine is approximately 4.5 kg and is considered moveable equipment. The Ivotronic Voting Machine is intended to be operated in an office environment. The Ivotronic Voting Machine is Class II equipment and is intended for use in a Pollution Degree 2 environment. Each unit has been evaluated to the requirements of UL 60950, Third Edition.

**ELECTRICAL RATINGS:**

120/ 240 Vac, 0.8A, 50 - 60 Hz - Ivotronic Voting Machine

**TEST PERFORMANCE**

A representative sample of the product was tested in accordance with the "Standard for Safety of Information Technology Equipment, Including Electrical Business Equipment," UL 60950; Third Edition.

The following tests were performed:

Description	Clause
Power Interface (Input) Test	1.6
Permanence of Markings Test	1.7.13
Accessibility Test	2.1.2
Mechanical Strength and Stress Relief	4.2.2, 4.2.3, 4.2.4 & 4.2.7
Heating Test	4.5
Electric Strength	5.2

Results of the tests indicate the specimen conforms to the applicable test criteria.

WYLE LABORATORIES, INC.  
Page 5

Report No.: 45827-02  
E S & S

Issued: 9/11/2001

**CONCLUSION**

A representative sample of the product covered by this report has been evaluated and found to comply with the applicable requirements of the "Standard for Safety of Information Technology Equipment," UL 60950, Third Edition.

STATE OF ALABAMA }  
COUNTY OF MADISON }

James R. Dearman being duly sworn, deposes and says: The information contained in this report is the result of complete and carefully conducted testing and is to the best of his knowledge true and correct in all respects.

*James R. Dearman*

SUBSCRIBED and sworn to before me this 11 day of Sept, 2001

*Thomas Carlock* **SEAL**  
Notary Public in and for the State of Alabama at Large

My Commission expires April 6, 2002

Wyle shall have no liability for damages of any kind to person or property, including special or consequential damages, resulting from Wyle's providing the services covered by this report.

TEST BY: *Barbara A. Brooks* 11 Sept 01  
Barbara A. Brooks, Lead Test Specialist Date

APPROVED BY: *Robert R. Loop* 11 Sep 01  
Robert R. Loop, Project Engineer Date

(jet)



Page No. L-8  
Test Report No. 45827-01

WYLE LABORATORIES, INC.  
Page 6

Report No.: 45827-02  
E S & S

Issued: 9/11/2001

**GENERAL INFORMATION:**

The Applicant and Manufacturer have agreed to produce and test Wyle tested products in accordance with the requirements of this report. The Manufacturer should notify Wyle to request authorization prior to using alternate parts, components, or materials.

**COMPONENTS:**

Components used shall be those shown in the Wyle report covering the products specified in the index including any amendments and/ or revisions.

**Symbol Authorization**

Based on the data presented in this report, the Ivotronic Voting Machine is authorized to display the Wyle Nationally Recognized Test Laboratory symbol below. The system should bear the symbol shown as evidence of compliance with the appropriate standard for safety.



**Listing File Number**

This product will be listed under Wyle Laboratories' File Number 45827 as long as the periodic site inspections demonstrate conformance to the mechanical and electrical configuration as delineated in this document. Revocation of the listing voids the authorization above.

**LISTING MARK:**

The Wyle listing mark applied to the products shall either be separable in form, such as labels purchased from Wyle Laboratories, or on a product nameplate or other media only as specifically authorized by Wyle Laboratories. Use of the listing mark is subject to the control of Wyle Laboratories.



WYLE LABORATORIES, INC.  
Page 7

Report No.: 45827-02  
E S & S

Issued: 9/11/2001

**MANUFACTURING AND PRODUCTION TESTS:**

Manufacturing and Production Tests shall be performed as required in this Report.

Electric Strength (High-potential) – Dielectric Strength at 1000 Vac (or Peak DC Equivalent) for one second.

**FOLLOW-UP SERVICE:**

Wyle Laboratories shall conduct random, quarterly, unannounced inspections to ensure conformance with the test and evaluation report, test standards, and field inspections, and to monitor and ensure proper use of the Wyle Product Safety Mark. Special attention will be given to the following:

1. Conformance of the manufactured product to the descriptions in this Report.
2. Conformance of the use of the Wyle mark with the requirements of this Report and the Listing, Labeling, and Follow-up Service Agreement.
3. In-plant quality control procedures and personnel.
4. Manufacturing changes.
5. Performance of specified Manufacturing and Production Tests.

In the event that the Wyle representative identifies non-conformance(s) to any provision of this Report, the Applicant shall take one or more of the following actions:

1. Correct the non-conformance.
2. Remove the Wyle Listing Mark from non-conforming product.
3. Contact the issuing product safety evaluation center for instructions.

**GENERAL REQUIREMENTS AND DEFINITIONS:**

**Recognized** – Identifies any component, part or subassembly covered under the recognition service of an NRTL (US) or a CO (Canada) and intended for use in Listed, Certified, or Recognized products.

**Listed** – Identifies any product covered under the Listing or Certification service of an NRTL or a CO.

Page No. L-10  
Test Report No. 45827-01

WYLE LABORATORIES, INC.  
Page 8

Report No.: 45827-02  
E S & S

Issued: 9/11/2001

**GENERAL REQUIREMENTS AND DEFINITIONS (Continued)**


**Construction Details** – For specific construction details, reference should be made to the following photographs and descriptions. All dimensions are approximate unless specified as exact or within a tolerance. In addition to the specific construction details described in this Report, the following general requirements also apply.

1. **Mechanical Assembly** – components such as switches and wiring terminals are reliably mounted and prevented from shifting or rotating by screws or the mounting format.
2. **Corrosion Protection** – All ferrous metal parts are suitably protected against corrosion by painting, plating or the equivalent.
3. **Internal Wiring** – Internal wiring is reliably routed away from sharp or moving parts. Internal wiring leads terminate in soldered connections made mechanically secure prior to soldering.
4. **Current Carrying Parts** – All current carrying parts are of silver, copper, or a copper base alloy.
5. **Accessibility of Live Parts** – All uninsulated live parts are housed within an enclosure and are adequately protected from contact by the articulated finger probe.
6. **Over-Voltage/Overload Protection** – The models are all protected against overvoltage by a overcurrent protection inherent in the power supply.
7. **Markings** – The unit is marked with the manufacturer's name, model number, electrical ratings, and cautionary markings where required.
8. **Instruction Manual** – An instruction manual is provided with each unit that is shipped from the factory. The instruction may be in the form of a separate booklet, or sheet, or may be part of the instruction manual, but in any case, they shall be separated in format from other instructions and shall appear before any operating instructions. The letters in text and illustrations in the instructions shall be clearly legible. "IMPORTANT SAFETY INSTRUCTIONS" and "SAVE THESE INSTRUCTIONS" shall be emphasized and clearly distinguishable from the rest of the text.
9. **Definitions** – Unless specifically stated otherwise, the following general definitions, terminology, and construction details apply:
  - a) **Dimensions** – All dimensions specified are approximate and are within plus or minus one-tenth of the base unit, unless stated otherwise.
  - b) **Listed** – Listed or certified by an accredited Certification Organization.
  - c) **Component** – Accepted by an accredited Certification Organization with certain restrictions, and appears in that organization's list of accepted components.
  - d) **Unlisted components** – No recognized third-party certification.

WYLE LABORATORIES, INC.  
Page 9

Report No.: 45827-02  
ES & S

Issued: 9/11/2001



**Figure 1**  
Ivotronic Voting Machine


**General** – Figure 1 shows the front and top view of the Ivotronic. The product measures approximately 50 mm high by 394 mm wide by 328 mm deep by 3 mm thick, with a weight of approximately 4.5 kg. The unit is considered moveable equipment.

1. Unit Enclosure (Bottom Cover) – Recognized Component Plastic (UL). ABS Resin, Grand Pacific Petrochemical Corp., D-100, approximately 3 mm thick, rated minimum 94HB. The bottom cover consists of the bottom and half of the sides, front and rear faces of the enclosure. The bottom cover measures approximately 25 mm high by 394 mm wide by 328 mm deep. Provided with various shaped and style cut-outs for connectors. Secured by screws.
2. Unit Enclosure (Top Cover) – Recognized Component Plastic (UL). ABS Resin, Grand Pacific Petrochemical Corp., D-100, approximately 3 mm thick, rated minimum 94HB. The top cover measures approximately 25 mm high by 394 mm wide by 328 mm deep. Provides with an opening measuring approximately 198 mm high by 256 mm wide for the LCD display. Secured by screws.

WYLE LABORATORIES, INC.  
Page 10

Report No.: 45827-02  
ES & S

Issued: 9/11/2001



**Figure 2**  
Ivotronic Voting Machine

**General** – Figure 2 shows the back and top view of the Ivotronic Voting Machine. The Ivotronic Voting Machine measures approximately 50 mm high by 394 mm wide by 328 mm deep by 3 mm thick, with a weight of approximately 4.5 kg. The unit is considered moveable equipment.

1. Unit Enclosure (Display Enclosure) – See Figure 1, Item 1 for details.
2. Unit Enclosure (Bottom Cover) – See Figure 1, Item 2 for details.

WYLE LABORATORIES, INC.  
Page 11

Report No.: 45827-02  
E S & S

Issued: 9/11/2001




Figure 3  
Power Source

**General** – Figure 3 shows the Ivotronic Voting Machine Power Supply.

1. **Power Source** – Located in Primary Circuit. Listed Component (UL). Group West, Part No. BUT-09-2770, Rated 9 Vdc, 2770 mA, 24.93 W. Provided with a non-detachable power cord, Type VW - 1SC. Provided with barrel-type connector at one end and a detachable cord set Type SPT-1, 18 AWG/ 2 C. Provided with non-industrial style NEMA 1-15P, attachment plug at one end rated 300 V, 15 A and non-industrial style IEC-320-C7, appliance connector at the other end rated 300 V, 15 A.

WYLE LABORATORIES, INC.  
Page 12

Report No.: 45827-02  
E S & S

Issued: 9/11/2001




Figure 4  
Printed Wiring Board and Battery Back-up

**General** – Figure 4 shows the printed wiring board and the DC Battery back-up.

1. Internal Wiring – Provided in SELV circuits. Suitable for the application and voltages applied. See General Requirements.
2. Lithium Battery – Provided in SELV circuit. Recognized Component (UL). Tadiran Lithium Batteries, Model Number TL-2155, Rated 3.6 Vdc, 1.45 Ah, 85°C. Secured by soldering.
3. Inverter – Provided in SELV circuit. TDK, Part Number CXA-M10A-L. Rated Input 6 Vdc, 1.5 A, 60°C, Rated Output 1200 Vac, 12A, 60°C. Secured by soldering.
4. Fuses – Provided in SELV circuit. Recognized Component (UL). Littelfuse, Part Number R459 003, Rated 125 V, 3A, 125°C. Secured by soldering.  
Alternate – Same as above except Bussman, Part Number SFT-3, Rated 125 V, 3A. Secured by soldering.
5. Back-up Battery Pack – Six provided in series. Provided in SELV circuit. Panasonic, Part Number HHR6550D, Rated 1.2 Vdc, 6800mAh, 60°C.
6. Printed Wiring Boards - Recognized Component (UL). Rated minimum 94-V1, secured by screws.

WYLE LABORATORIES, INC.  
Page 13

Report No.: 45827-02  
E S & S

Issued: 9/11/2001




Figure 5  
IvoTronic Machine and Floor Stand

**General** – Figure 5 shows the IvoTronic Machine and Floor Stand. The Floor Stand measures approximately 147 mm high by 614 mm wide by 462 mm deep by 5.9 mm thick, with a weight of approximately 7.2 kg. The unit is considered moveable equipment.

1. **Enclosure (Top)** – Recognized Component Plastic (UL). ABS Resin, Grand Pacific Petrochemical Corp., D-100, approximately 3 mm thick, rated minimum 94HB. The top cover consists of the top and half of the sides, front and rear faces of the enclosure. The top cover measures approximately 72 mm high by 614 mm wide by 462 mm deep. Secured by screws.
2. **Enclosure (Bottom)** – Recognized Component Plastic (UL). ABS Resin, Grand Pacific Petrochemical Corp., D-100, approximately 3 mm thick, rated minimum 94HB. The bottom cover consists of the bottom and half of the sides, front and rear faces of the enclosure. The bottom cover measures approximately 72 mm high by 614 mm wide by 462 mm deep. Provided with oblong, angled openings measuring approximately 67 mm in diameter by 64 mm long for Item 3. Secured by screws.
3. **Leg** – Four provided. Brushed of Aluminum, approximately 1.9 mm thick. Each leg measures approximately 254 mm in diameter by 1002 mm long. Secured by rivets.

Report No.: 45827-02  
E S & S

WYLE LABORATORIES, INC.  
Page A-1

Issued: 9/11/2001

**ATTACHMENT A  
USER SAFETY WARNINGS**

WYLE LABORATORIES, INC.  
Page A-2

Report No.: 45827-02  
E S & S

Issued: 9/11/2001

6

## Chapter 2: Warning Messages

When used properly, the iVotronic is safe and effective. The following symbols are used throughout this manual to indicate when hazards may occur during normal operation of the iVotronic. Please read the warnings and proceed with caution when carrying out iVotronic operations that are potentially hazardous.



**Electrical Shock Danger:** This symbol indicates a danger of electric shock. There are high voltages present inside the enclosure of the iVotronic. To reduce the risk of fire or electric shock, do not attempt to open the enclosure or gain access to areas for which you have not received training. Only ES&S qualified personnel should open the enclosure of the iVotronic.



**User Caution:** This user caution symbol indicates that damage to the iVotronic or injury to the user could occur if the proper procedures are not followed. Carefully follow all instructions and proceed with caution when this symbol is associated with a set of instructions.

## Important Safety Instructions



**Important:** Carefully read and follow these safety instructions.

### Read the User's Guide

Read the user's guide carefully before operating the iVotronic. Carefully follow all instructions and read all warnings.

### Power Sources

This product should be operated only from the type of power source indicated in the ID label. Installation should be in compliance with applicable sections of the National Electric Code. Consult your local building code before installing.

Page No. L-18  
Test Report No. 45827-01

---

Report No.: 45827-02  
E S & S

WYLE LABORATORIES, INC.  
Page A-3

Issued: 9/11/2001

8

**Water and Moisture**



**Caution:** Do not place containers with liquids such as coffee, water or soda on or around the iVotronic. Do not operate the iVotronic in an excessively wet environment. Store the iVotronic in a cool dry place.

**Cleaning**

Follow instructions in the operator's manual for cleaning the iVotronic. Only use cleaning solutions approved by ES&S in the operator's manual.

**Heating**

Do not install this product near heat sources such as radiators, air ducts, areas subject to direct sunlight or other products that produce excessive heat.

**Power Cord Protection**

The power supply cord for this product should be routed or installed in such a manner to protect it from being walked over or pinched. The unit should be powered down completely before connecting or disconnecting the power cord. The power cord should be removed before moving the unit. Only plug the power cord in to an easily accessible unobstructed wall socket.

**Servicing**

Do not attempt to service the iVotronic unless specifically instructed to do so by ES&S. Do not attempt to gain access to areas of the unit where dangerous voltages are present. Only qualified ES&S technicians should service the iVotronic.

WYLE LABORATORIES, INC.  
Page A-4

Report No.: 45827-02  
ES & S

Issued: 9/11/2001

---

7

#### Damage Requiring Service

Unplug the iVotronic and call ES&S to consult a qualified service technician under the following conditions:

- When the power cord is damaged.
- If liquid has been spilled into the iVotronic casing.
- Consult a technician if the iVotronic does not function normally while following instructions in the operator's manual. Adjust only the controls specified in the operator's manual. Improper adjustment of other controls may result in damage to the iVotronic and will often require work by an ES&S technician to restore the iVotronic to normal operating condition.
- If the product is damaged in any way.
- When the iVotronic displays a negative change in performance.

#### Battery Replacement Warning



Caution: Danger of explosion if battery is incorrectly replaced. Replace only with same or equivalent type battery recommended by ES&S. Dispose of used batteries properly.

Page No. L-20  
Test Report No. 45827-01

---

WYLE LABORATORIES, INC.  
Page B-1

Report No.: 45827-02  
E S & S

Issued: 9/11/2001

**ATTACHMENT B**  
**TEST DATA**

Page No. L-21  
 Test Report No. 45827-01

WYLE LABORATORIES, INC.  
 Page B-2

Report No.: 45827-02  
 E S & S

Issued: 9/11/2001

Job No.: 45827 Date: 5/24/2001  
 Specimen ID: IVO tronic S/N: Y0010062  
 Customer: ES&S



Power Interface (Input) Test Data Sheet		Clause	CSA 60950 (00)/UL 60950	EN 60950
		1.6		1.6
Acceptance Criteria (or Maximum Allowable Limits):		Maximum 10% Deviation		
Compliance: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		Tested by: <i>[Signature]</i> Date: 5/24/01		
		Approved by: <i>[Signature]</i> Date: 22 Aug 01		
Test No.	Input			
	Volts	Hz	Amps	Watts
1	120	60	.116 mA	13.92
2				
3				
4				
5				
6				
Test Conditions				
1	Power supply is a listed power source			
2				
3				
4				
5				
6				

\*For power supply outputs and/or convenience outlet receptacle.

WJL-1481, Rev. May '01

(10)

Page No. L-22  
 Test Report No. 45827-01

WYLE LABORATORIES, INC.  
 Page B-3

Report No.: 45827-02  
 E S & S

Issued: 9/11/2001

Job No.: 45827 Date: 5/24/2001  
 Specimen ID: IVO tronic S/N: V0010062  
 Customer: ES&S



Permanence of Marking Test	Standard	CSA 60950 (00)/UL 60950	EN 60950	EN 61010-1
	Clause	1.7.13	1.7.13	5.3
Compliance: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		Tested by: <i>Janet Stone</i> Date: 5/24/01 Approved by: <i>Samuel Brink</i> Date: 5/24/01		
<b>Method:</b> A sample of the marking label was subjected to the following test. The surface of each marking as noted below was rubbed by hand for a period of 15 seconds with a water soaked cloth and again for a period of 15 seconds with a petroleum spirit (Hexane) soaked cloth. The results are listed below.				
Test Conditions				
Use of marking?	Compliant			
Material?	Mylar			
Held By?	Pressure sensitive adhesive			
Applied surface material	Plastic Enclosure			
Observations:	Water	Hexane		
Any Damage?	NO	NO		
Legible?	Yes	Yes		
Curled?	NO	NO		
Edges Lifted?	Compliant	Compliant		
Easily removed intact?	NO	NO		
Comments:				

WH-1528, Rev. May '01

(tab)

WYLE LABORATORIES, INC.  
 Page B-4

Report No.: 45827-02  
 ES & S

Issued: 9/11/2001

Job No.: 45827 Date: 5/24/2001  
 Specimen ID: IVO tronic S/N: V0010062  
 Customer: ES&S



Accessibility Test	Standard	CSA 60950 (00)/UL 60950	EN 60950
	Clause(a)	2.1.2 & 2.8.3	2.1.2 & 2.8.3
Acceptance Criteria (for Maximum Allowable Limits):			
Compliance: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Tested by: <u>Annalyn Jones</u> Date: <u>5/24/01</u> Approved by: <u>Richard Brock</u> Date: <u>5/24/01</u>		
The EUT with all operator access covers removed was subjected to this test. A test finger was applied without appreciable force to all apertures, in an attempt to contact hazardous parts. Operator detachable connectors were tested in an attempt to contact hazardous parts. Operator detachable connectors were tested during and after disconnections. Openings preventing the entry of the test finger were further tested by means of a straight unjointed version of the test finger, which was applied with a force of 30 N (6.75 lbs). If entry of the unjointed finger was possible, the test with the articulated test finger was repeated with the finger being pushed through the aperture, if necessary.			
Results: Compliant No operator accessible parts			
The EUT with all operator detachable parts, including fuse holders and lamps were left in place and operator access doors and covers closed, was subjected to this test. A test pin was applied to all apertures located in electrical enclosures, in an attempt to contact hazardous parts.			
Results: Compliant No hazardous accessible parts/ voltages			
The EUT employing a safety interlock was subjected to this test. The test finger was applied to all covers, guards, doors, etc., to determine that if inadvertent reactivation of the interlock circuit did occur.			
Results: Compliant No accessible hazardous parts/ voltages			

WH-1529, Rev. May '01

(66)

WYLE LABORATORIES, INC.  
Page B-5

Report No.: 45827-02  
ES & S

Issued: 9/11/2001



## MECHANICAL STRENGTH AND STRESS RELIEF

Job No.: 45827 Date: 5/25/2001  
Specimen ID: IVQtronic S/N: V0010062  
Customer: ES&S

These measurements were taken in accordance with the following standards:

Underwriters Laboratories 60950;  Canadian Standard C22.2 #60950;  European Standard EN 60950;  
 European Standard 61010-1;  Other: \_\_\_\_\_

### Section 4.2.2 - Steady Force 10N

A steady force of  $10N \pm 1N$  was applied to components and parts, other than parts serving as an Enclosure.

Results: Compliant

### Section 4.2.3 - Steady Force 30N

The EUT was placed on a suitable surface. A steady force of  $30N \pm 3N$  was applied for a period of 5 seconds on four surfaces by means of a straight unjointed test finger.

Results: Compliant

### Section 4.2.4 - Steady Force 250N

The EUT was placed on a suitable surface. A steady force of  $250N \pm 10N$  was applied for a period of 5 seconds on four surfaces by means of a suitable tool which provided contact over a circular plane surface 30 mm in diameter.

Results: Compliant

### Section 4.2.5 - Steel Ball Test

With the sample held in a fixed position, a smooth sphere, approximately 50 mm in diameter and weighing  $500g \pm 25g$ , was allowed to fall horizontally from rest through the distance, 1300 mm required to cause the sphere to strike the sample in three different locations.

Results: Compliant

### Section 4.2.7 - Mold Stress Relief

A sample consisting of the complete equipment, or of the complete enclosure, together with any supporting framework, is subjected to a circulating air oven to a temperature 10 K higher than the maximum temperature observed during the test of 5.1, but not less than 70 °C, for a period of 7 hours, then permitted to cool to room temperature.

Results: Compliant

Technician: Jennifer Thomas Date: 5/25/01  
Engineer: Richard Brooks Date: 30 Aug 01

WH-1519, Rev. May '01

Sheet No. 1 of 1



WYLE LABORATORIES, INC.  
 Page B-6

Report No.: 45827-02  
 ES & S

Issued: 9/11/2001



**NORMAL TEMPERATURE  
 TEST DATA SHEET**

Job No.: 45827 Date: 6/13/01  
 Specimen ID: I/Otronic S/N: V0010062  
 Customer: ES & S

These measurements were taken in accordance with the following standards:

- Underwriters Laboratories 60950;  Canadian Standard C22.2 #60950;  European Standard EN 60950;  
 European Standard 61010-1;  Other: \_\_\_\_\_

TC No.	TC Location	Readings <input checked="" type="checkbox"/> °C <input type="checkbox"/> °F				
1.	T-1	56.3				
2.	U-8	43.2				
3.	L-10	47.0				
4.	C-1	43.8				
5.	U-4	34.7				
6.	U-15	30.6				
7.	U-1	33.2				
8.	Battery	29.0				
9.	U-1 Epson	36.8				
10.	Battery Pk	49.1				

Voltage Input to E.U.T.: 120 VAC

Ambient Temperature: 24°C Relative Humidity: 43 %

Thermocouple Meter: 115683

Notes on Data: \_\_\_\_\_

Remarks: \_\_\_\_\_

Technician: [Signature] Date: 6/13/01

Engineer: [Signature] Date: 29 Aug 01

WH-1437 Rev. May 01

Sheet No. 1 of 2

Page No. L-26  
 Test Report No. 45827-01

WYLE LABORATORIES, INC.  
 Page B-7

Report No.: 45827-02  
 ES & S

Issued: 9/11/2001



**NORMAL TEMPERATURE  
 TEST DATA SHEET**

Job No.: 45827 Date: 6/13/2001  
 Specimen ID: IVOtronic Acces box S/N: V0010062  
 Customer: ES&S

These measurements were taken in accordance with the following standards:

- Underwriters Laboratories 60950;  Canadian Standard C22.2 #60950;  European Standard EN 60950;  
 European Standard 61010-1;  Other: \_\_\_\_\_

TC No.	TC Location	Readings <input checked="" type="checkbox"/> °C / <input type="checkbox"/> °F			
1.	Not Used	N/A			
2.	F- 2	26.3			
3.	U- 2	26.2			
4.	U- 3	26.3			
5.	Not Used	N/A			
6.	Not Used	N/A			
7.	Not Used	N/A			
8.	Not Used	N/A			
9.	Not Used	N/A			
10.	C- 4	26.1			

Voltage Input to E.U.T.: 120 VAC

Ambient Temperature: 24°C Relative Humidity: 43%

Thermocouple Meter: 115683

Notes on Data: \_\_\_\_\_

Remarks: \_\_\_\_\_

Technician: *Jennifer Thompson* Date: 6/13/01

Engineer: *Richard A. Brooks* Date: 30 Aug 01

WH-1437 Rev. May 01

Sheet No. 2 of 2



Page No. L-27  
 Test Report No. 45827-01

WYLE LABORATORIES, INC.  
 Page B-8

Report No.: 45827-02  
 ES & S

Issued: 9/11/2001

Job No.: 45827  
 Specimen ID: I/Otronic  
 Customer: ES&S

Date: 5/24/2001  
 S/N: V0010062



Electric Strength Test Data Sheet	Standard	CSA 60950(00)/UL 60950	EN 60950
	Clause	5.2	5.2
Acceptance Criteria (or Maximum Allowable Limits): No insulation breakdown during test.			
Compliance: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		Tested by <u>Annal Storer</u> Date: <u>5/24/01</u> Approved by <u>Robert W. Work</u> Date: <u>2/28/01</u>	
Location	Test Voltage	Result	
(A) ON COMPLETE SYSTEM:	1414	Pass	
Primary and Earth			
Primary and SELV Secondary			
Primary and SELV Secondary			
Primary and ___ V Secondary			
Primary and ___ V Secondary			
SELV Secondary and Earth			
SELV Secondary and Earth			
___ V Secondary and Earth			
___ V Secondary and Earth			
(B) ON SAFETY ISOLATING TRANSFORMER:			
Primary and SELV Secondary			
Primary and Core/Screen			
Primary and ELV Secondary			
Secondary and Core			
Between SELV and Secondaries			
Between ELV and SELV Secondaries			
(C) ON TNV CIRCUITS: (including alternative optical isolators and relays)			
Telephone Ringing Circuit and Earth			
Telephone Ringing Circuit and Secondary Circuits			
Comments:			

WH-1473, Rev. May 01

610

Report No.: 45827-02  
E S & S

WYLE LABORATORIES, INC.  
Page C-1

Issued: 9/11/2001

**ATTACHMENT C**  
**INSTRUMENTATION EQUIPMENT SHEET**

WYLE LABORATORIES, INC.  
Page C-2

Report No.: 45827-02  
E S & S

Issued: 9/11/2001



INSTRUMENTATION EQUIPMENT SHEET

DATE: 6/12/01      JOB NUMBER: 45827      TEST AREA: PRODUCT SA  
TECHNICIAN: J. STONER      CUSTOMER: ES&S      TYPE TEST: UL 60950

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	SCALE	SETRA	SUPER COUNT	15864	113735	27LBS	±.0005LBS	4/25/01	4/25/02

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION

*Jennifer Stoner 6/12/01*

CHECKED & RECEIVED BY

*Bonnie B. ...*

Q.A.

*Ronda ...*

WHL-1029A, REV. A7E 79

This page intentionally left blank.



7800 Highway 20 West  
Huntsville, Alabama 35806  
Phone (256) 837-4411 Fax (256) 830-2109  
www.wylelabs.com

REPORT NO.: 46229-01  
WYLE JOB NO.: 46229  
CLIENT P.O. NO.: 720556  
CONTRACT: N/A  
TOTAL PAGES (INCLUDING COVER): 16  
DATE: September 19, 2001

## TEST REPORT

47 CFR PART 15, SUBPART B,  
CLASS B TESTING  
ON THE  
IVOTRONIC VOTING MACHINE  
(15-INCH VERSION)

For  
Pivot International  
14125 West 95<sup>th</sup> Street  
Lenexa, KS 66215

STATE OF ALABAMA  
COUNTY OF MADISON

James R. Dearman, being duly sworn, deposes and says: The information contained in this report is the result of complete and carefully conducted testing and is to the best of his knowledge true and correct in all respects.

SUBSCRIBED and sworn to before me this 21 day of Sept 2001

Thomas Landlock SEAL  
Notary Public in and for the State of Alabama at Large

My Commission expires April 6 2007

Wyle shall have no liability for damages of any kind to person or property, including special or consequential damages, resulting from Wyle's providing the services covered by this report.

TESTED BY: Tim Yancy 21 Sept 01  
Tim Yancy, Test Specialist Date

REVIEWED BY: Barbara Brooks 21 Sept 01  
Barbara Brooks, Lead Test Specialist Date

(dsc)



COPYRIGHT BY WYLE LABORATORIES. THE RIGHT TO REPRODUCE, COPY, EXHIBIT, OR OTHERWISE UTILIZE ANY OF THE MATERIAL CONTAINED HEREIN WITHOUT THE EXPRESS PRIOR PERMISSION OF WYLE LABORATORIES IS PROHIBITED. THE ACCEPTANCE OF A PURCHASE ORDER IN CONNECTION WITH THE MATERIAL CONTAINED HEREIN SHALL BE EQUIVALENT TO EXPRESS PRIOR PERMISSION.  
WH-1404, Rev. Feb '97

**1.0 INTRODUCTION**

**1.1 Scope**

This report documents conformance with the FCC Rules, Part 15, as listed below, and details the results of the testing performed on one Pivot International iVotronic Voting Machine (15-inch version), at Wyle Laboratories' Huntsville, Alabama, Test Facility. The specimen was received in good condition on August 13, 2001

**1.2 Product Description**

The iVotronic Voting Machine is utilized for collecting and tallying votes. The iVotronic is provided with a 15-inch LCD.

A detailed description of the test specimen (iVotronic) and the support equipment is provided in Section 2.1.

**1.3 References**

- Pivot International's Purchase Order No. 720556
- Wyle Laboratories' Request for Quotation No. 545/011638
- ANSI C63.4-1992, "Methods of Measurement of Radio Noise From Low-Voltage Electrical and Electronic Equipment in the Range of 9 kHz to 40 GHz"
- Code of Federal Regulations (CFR) Title 47, "Telecommunications"
- Wyle Laboratories' Quality Assurance Program Manual, Revision 1
- ANSI/NCSL Z540-1, "Calibration Laboratories and Measuring and Test Equipment, General Requirements"
- ISO 10012-1, "Quality Assurance Requirements for Measuring Equipment"
- MIL-STD-45662A, "Calibration System Requirements"

**1.0 INTRODUCTION (Continued)**

**1.4 Quality Assurance**

**1.4.1 Quality Assurance Program**

All work performed on this test program was completed in accordance with Wyle Laboratories' Quality Assurance Program. The Wyle Laboratories, Huntsville Facility, Quality Management System is registered in compliance with the ISO-9001 International Quality Standard. Registration has been completed by Quality Management Institute (QMI), a Division of Canadian Standards Association (CSA).

**1.4.2 Test Equipment and Instrumentation**

All instrumentation, measuring, and test equipment used in the performance of this test program were calibrated in accordance with Wyle Laboratories' Quality Assurance Program, which complies with the requirements of ANSI/NCCL Z540-1, ISO 10012-1, and Military Specification MIL-STD-45662A. Standards used in performing all calibrations are traceable to the National Institute of Standards and Technology (NIST) by report number and date. When no national standards exist, the standards are traceable to international standards or the basis for calibration is otherwise documented.

Attachment C contains a list of the instrumentation, measuring, and test equipment that was used to perform the tests.

**2.0 SYSTEM TEST CONFIGURATION**

**2.1 Details of Tested System**

The identifiers for the test specimen, support devices, and cables used in the tested system are:

• **Test Specimen**

Item	Part No.	Serial No.	Quantity
iVotronic	15-inch Version	V0100690A	1
Auto Vote Cartridge	Not Provided	PV 0101717-A	1
Power Supply	BUT-09-2770	001*	1

\*Assigned by Wyle Laboratories.

• **Support Test Equipment**

No support equipment was used in the testing of the iVotronic.

**2.2 Input Power**


The FCC Part 15 testing was performed using 120 VAC/60 Hz input power.

**2.0 SYSTEM TEST CONFIGURATION (Continued)**

**2.3 Modifications**

No modifications were required to achieve compliance.

**2.4 iVotronic Test Configuration**



**3.0 EMISSIONS TESTING**

**3.1 Test Facility**

The Open-Area Test Site facility used to collect the data in this report is located on the grounds of Wyle Laboratories' Huntsville, Alabama Test Facility. This site is fully described in reports provided to the FCC, Commission Reference 31040/SIT1300B3. This site was tested and complies with the requirements of ANSI C63.4-1992.

**3.0 EMISSIONS TESTING**

**3.2 Test Procedure**

Radiated emissions from the iVotronic shall be measured in accordance with the procedure outlined in ANSI C63.4-1992. Compliance of the test specimen shall be based on comparison with the Class B limits that are specified in FCC Rules, Part 15. The test shall determine radiated emissions levels from 30 MHz to 1000 MHz and be obtained at an antenna-to-test specimen distance of 3 meters. The iVotronic shall not generate radiated emissions that exceed these specified limits.

Conducted emissions from the iVotronic shall be measured in accordance with the procedure outlined in FCC Rules, Part 15. Compliance of the iVotronic will be based on comparison with the limits that are specified above. The iVotronic shall not generate conducted emissions that exceed these specified limits.

3.0 EMISSIONS TESTING (Continued)

3.2 Test Procedure (Continued)

3.2.1 Field Strength Calculations

The field strength emissions are calculated by adding the Antenna Factor and the Cable Factor and subtracting the Amplifier Gain (if any) from the measured reading. The basic equation with a sample calculation is as follows.

$$FS = RA + AF + CF - AG + DC$$

where:  $FS$  = Field Strength  
 $RA$  = Received Amplitude  
 $AF$  = Antenna Factor  
 $CF$  = Cable Attenuation Factor  
 $AG$  = Amplifier Gain  
 $DC$  = Distance Correction

For example: assume a receiver input reading of 52.5 dB $\mu$ V is obtained. The Antenna Factor, 7.4 dB, and Cable Factor, 1.1 dB, are added and the amplifier gain, 29 dB, is subtracted.

$$FS = 52.5 + 7.4 + 1.1 - 29$$

$$FS = 32.0 \text{ dB}\mu\text{V/m}$$


Correction for measurement distance from 3 meters to 30 meters (if necessary) would add the distance correction factor of -20.0 dB to the value of  $FS$ .

$$FS (3 \text{ meters}) = 32.0 \text{ dB}\mu\text{V/m} - 20.0$$


$$FS (30 \text{ meters}) = 12.0 \text{ dB}\mu\text{V/m}$$

3.3 Test Results

The radiated and conducted emissions from the Ivotronic complied with FCC Part 15, Class B, limits. No anomalies were noted. The test data is presented in Attachment B.



**Photograph A-1**  
**Radiated Emissions Test Setup**



**Photograph A-2**  
**Conducted Emissions Test Setup**

---

**WYLE LABORATORIES**  
Huntsville Facility



Wyle Labs  
 7800 Highway 20 West  
 258-837-4411

Customer: PMOT  
 Specification: FCC B RADIATED

Work Order #	46229	Date	Mon Aug-13-2001
Test Type	Radiated Scan	Time	09:54:34
Equipment	VOTING MACHINE	Sequence	0
Manufacturer	ES&S	Tested By	T.Yancy
Model	1700		
S/N	V0100690A		

Equipment Under Test (* = EUT):			
Function	Manufacturer	Model	S/N
VOTING	ES&S	1700	V0100690A


Support Devices:			
Function	Manufacturer	Model	S/N
None			

Test Conditions / Notes:  
 ACTIVE VERTICAL HORIZONTAL

Transducer Legend:	
T1=3m cables	T0=3meter ant.fac


Measurement Data:		Readings listed by frequency.				Test Distance: 3 Meters				
#	Freq MHz	Band dBµV	31	32	Dist dBµV/m	Corr dBµV/m	Spec dBµV/m	Power	Type	Margin
1	69.520	27.5	+1.0	+8.4	+0.0	37.9	40.0	Vert	Quasi Peak	-2.1
2	72.711	27.7	+1.1	+8.9	+0.0	37.7	40.0	Vert	Quasi Peak	-2.3
3	78.984	28.8	+1.1	+11.6	+0.0	36.4	40.0	Vert	Quasi Peak	-1.8
4	240.031	20.8	+2.0	+11.6	+0.0	34.2	46.0	Horiz	Quasi Peak	-11.8
5	280.029	22.1	+2.1	+12.7	+0.0	36.9	46.0	Horiz	Quasi Peak	-8.4
6	300.026	18.0	+2.2	+13.2	+0.0	34.4	46.0	Horiz	Quasi Peak	-11.8
7	326.033	22.9	+2.3	+13.7	+0.0	36.9	46.0	Horiz	Quasi Peak	-7.1
8	378.024	19.3	+2.4	+16.1	+0.0	36.8	46.0	Horiz	Quasi Peak	-8.2

*Barbara Brooks 8/13/01*  
*Barbara Brooks 6 Sept 01*




WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility




WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility




WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility




WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility




WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility




WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility



WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility



WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility



INSTRUMENTATION EQUIPMENT SHEET

DATE:	8/14/01	JOB NUMBER:	46229	TEST AREA:	FCC SITE				
TECHNICIAN:	T.YANCY	CUSTOMER:	PIVOT INT.	TYPE TEST:	EMISSIONS				
NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	Q-PEAK ADAPTER	HP	85650A	2811A01189	112109	BY PASS MOD	3dB	5/22/01	5/22/02
2	SPEC ANAL	HP	8566B	2437A03750	R03750	MFO	CBRT	6/ 6/01	6/ 6/02
3	PRSELECTOR	HP	85685A	3107A01286	109934	20-20HZ	2dB	9/21/00	9/21/01
4	LISN	SOLAR	8028-30-TS-24-B	974622	113973	10K-100MHZ	CBRT	10/ 6/00	10/ 6/01
5	ANTENNA	EMCO	EM-6917A	124116	114415	30MHZ - 30HZ	CBRT	5/22/01	5/22/02
6	LISN	SOLAR	8028-30-TS-24-B	974623	113974	10K-100MHZ	CBRT	10/ 6/00	10/ 6/01

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION

*[Signature]* 8-8-01 CHECKED & RECEIVED BY *[Signature]* 8/15/01  
*[Signature]* 8/15/01

WS-109A, REV. APR 99

WYLE LABORATORIES  
 Huntsville Facility

**wyle**  
laboratories

7800 Highway 20 West  
Huntsville, Alabama 35806  
Phone (256) 837-4411 Fax (256) 830-2109  
www.wylelabs.com

REPORT NO.: 46229-02  
WYLE JOB NO.: 46229  
CLIENT P.O. NO.: 720556  
CONTRACT: N/A  
TOTAL PAGES (INCLUDING COVER): 12  
DATE: September 19, 2001

## TEST REPORT

47 CFR PART 15, SUBPART B,  
CLASS B TESTING  
ON THE  
IVOTRONIC VOTING MACHINE  
(12-INCH VERSION)


For  
Pivot International  
14125 West 95<sup>th</sup> Street.  
Lenexa, KS 66215

STATE OF ALABAMA  
COUNTY OF MADISON

James R. Dearman being duly sworn, deposes and says: The information contained in this report is the result of complete and carefully conducted testing and is to the best of his knowledge true and correct in all respects.



SUBSCRIBED and sworn to before me this 21 day of Sept 2001

 **SEAL**

Notary Public for the State of Alabama at Large

My Commission expires April 6, 2002

Wyle shall have no liability for damages of any kind to person or property, including special or consequential damages, resulting from Wyle's providing the services covered by this report.

TESTED BY: Tim Yancy 21 Sept 01  
Tim Yancy, Test Specialist Date

REVIEWED BY: Barbara A. Brooks 21 Sept 01  
Barbara Brooks, Lead Test Specialist Date

(dsc)



COPYRIGHT BY WYLE LABORATORIES. THE RIGHT TO REPRODUCE, COPY, EXHIBIT, OR OTHERWISE UTILIZE ANY OF THE MATERIAL CONTAINED HEREIN WITHOUT THE EXPRESS PRIOR PERMISSION OF WYLE LABORATORIES IS PROHIBITED. THE ACCEPTANCE OF A PURCHASE ORDER IN CONNECTION WITH THE MATERIAL CONTAINED HEREIN SHALL BE EQUIVALENT TO EXPRESS PRIOR PERMISSION.

WH-1404, Rev. Feb '97

**1.0 INTRODUCTION**

**1.1 Scope**

This report documents conformance with the FCC Rules, Part 15, as listed below, and details the results of the testing performed on one Pivot International iVotronic Voting Machine (12-inch version), at Wyle Laboratories' Huntsville, Alabama, Test Facility. The specimen was received in good condition on September 5, 2001.

**1.2 Product Description**

The iVotronic Voting Machine is utilized for collecting and tallying votes. The iVotronic is provided with a 12-inch LCD.

A detailed description of the test specimens (iVotronic) and the support equipment is provided in Section 2.1.

**1.3 References**

- Pivot International's Purchase Order No. 720556
- Wyle Laboratories' Request for Quotation No. 545/011638
- ANSI C63.4-1992, "Methods of Measurement of Radio Noise From Low-Voltage Electrical and Electronic Equipment in the Range of 9 kHz to 40 GHz"
- Code of Federal Regulations (CFR) Title 47, "Telecommunications"
- Wyle Laboratories' Quality Assurance Program Manual, Revision 1
- ANSI/NCSL Z540-1, "Calibration Laboratories and Measuring and Test Equipment, General Requirements"
- ISO 10012-1, "Quality Assurance Requirements for Measuring Equipment"
- MIL-STD-45662A, "Calibration System Requirements"

1.0 INTRODUCTION (Continued)

1.4 Quality Assurance

1.4.1 Quality Assurance Program

All work performed on this test program was completed in accordance with Wyle Laboratories' Quality Assurance Program. The Wyle Laboratories, Huntsville Facility, Quality Management System is registered in compliance with the ISO-9001 International Quality Standard. Registration has been completed by Quality Management Institute (QMI), a Division of Canadian Standards Association (CSA).

1.4.2 Test Equipment and Instrumentation

All instrumentation, measuring, and test equipment used in the performance of this test program were calibrated in accordance with Wyle Laboratories' Quality Assurance Program, which complies with the requirements of ANSI/NCCL Z540-1, ISO 10012-1, and Military Specification MIL-STD-45662A. Standards used in performing all calibrations are traceable to the National Institute of Standards and Technology (NIST) by report number and date. When no national standards exist, the standards are traceable to international standards or the basis for calibration is otherwise documented.

Attachment C contains a list of the instrumentation, measuring, and test equipment that was used to perform the tests.

2.0 SYSTEM TEST CONFIGURATION

2.1 Details of Tested System

The identifiers for the test specimen, support devices, and cables used in the tested system are:

• Test Specimen

Item	Part No.	Serial No.	Quantity
iVotronic	12-inch version.	100243	1
Auto Vote Cartridge	Not Provided	PV 0101717-A	1
Power Supply	BUT-09-2770	001*	1



\*Assigned by Wyle Laboratories.

• Support Test Equipment



No support equipment was used in the testing of the iVotronic.

2.2 Input Power

The FCC Part 15 testing was performed using 120 VAC/60 Hz input power.





WYLE LABORATORIES  
Huntsville Facility





WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility



WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility



WYLE LABORATORIES  
Huntsville Facility

WYLE LABORATORIES  
Huntsville Facility



INSTRUMENTATION EQUIPMENT SHEET

NO.	INSTRUMENT	MANUFACTURER	MODEL #	SERIAL #	WYLE #	RANGE	ACCURACY	CAL DATE	CAL DUE
1	Q-PEAK ADAPTER	HP	85650A	2811A01189	112109	BY PASS MOD	3db	5/22/01	5/22/02
2	SPBC ANAL	HP	8566B	2637A03750	R03750	MFO	CERT	6/ 6/01	6/ 6/02
3	PRESELECTOR	HP	85685A	3107A01286	109934	20 - 20GHZ	2db	9/21/00	9/21/01
4	LISN	SOLAR	8028-S0-TS-24-B	974632	113973	10K-100MHZ	CERT	10/ 6/00	10/ 6/01
5	ANTENNA	BMCO	BM-6917A	124116	114415	30MHZ - 30HZ	CERT	5/22/01	5/22/02
6	LISN	SOLAR	8028-S0-TS-24-B	974623	113974	10K-100MHZ	CERT	10/ 6/00	10/ 6/01

DATE: 8/14/01      JOB NUMBER: 46229      TEST AREA: FCC SITE  
 TECHNICIAN: T.YANCY      CUSTOMER: PIVOT INT.      TYPH TEST: EMISSIONS

This is to certify that the above instruments were calibrated using state-of-the-art techniques with standards whose calibration is traceable to the National Institute of Standards and Technology.

INSTRUMENTATION 8-8-01 CHECKED & RECEIVED BY Andrew F. [Signature] 8/15/01  
[Signature] 8/15/01

WR-1079A, REV. APR 79

WYLE LABORATORIES  
 Huntsville Facility

Page No. M-31  
Test Report No. 45827-01

---

Page No. C-2  
Test Report No. 46229-02

---

WYLE LABORATORIES  
Huntsville Facility

---

WYLE LABORATORIES  
Huntsville Facility

---

**This page intentionally left blank.**